

Ultra-Wideband RF Front-End for Real-Time Brain Signal Acquisition in Neuroprosthetic Devices

A. Suresh kumar^{1*}, J Krishnamoorthy²

¹Assistant Professor, Department of Computer Science and Engineering, Excel Engineering College Namakkal

²Assistant Professor, Dept. of information technology, Saveetha Engineering College, Saveetha Nagar, Thandalam, Chennai, India

KEYWORDS:

Ultra-wideband (UWB),
RF front-end,
Brain signal acquisition,
Neuroprosthetic devices,
Low-noise amplifier (LNA),
Real-time neural recording,
Implantable electronics,
Biomedical telemetry,
CMOS design,
Wireless brain-machine interface (BMI).

ARTICLE HISTORY:

Submitted : 10.08.2025
Revised : 22.09.2025
Accepted : 19.12.2025

<https://doi.org/10.17051/NJRFCS/03.02.05>

ABSTRACT

A high-performance ultra-wideband (UWB) radio frequency (RF) front-end architecture that is inherently designed with a capability to acquire brain signals in real-time in next-generation neuroprosthetic systems is presented, developed, and experimentally validated in this paper. As more wireless brain machine interfaces (BMI) and closed-loop neuroprosthetic devices become clinically useful, the RF front-end circuitry has been a critical need in support of transmission of low-amplitude, high-resolution neural activity; such as intracranial electroencephalography (iEEG) and local field potentials (LFPs) with minimal delays, high fidelity and ultra-low power requirements. The suggested front-end consists of an integrated-global positioning system (GPS) front-end architecture that is miniaturized and fully integrated including a low-noise amplifier (LNA), high-Q bandpass filter, passively conceived frequency mixer with low-phase-noise local oscillator (LO) and high-rate analog-to-digital conversion interface. The system is designed based on sophisticated 65 nm CMOS production technology and is working effectively at an operating frequency of up to 0.1-6GHz UWB and optimized in terms of implantable or wearable models. Front-end circuit-level Simulation and analysis range conjugated with Cadence and ADS determine a gain of 18.5 dB, noise figure of less than 1.2 dB and a 2.3 mW power consumption with an end-to-end latency of less than 5 microseconds, assured within the safe limits of biomedical applications. Co-simulation of electromagnetic in CST also verifies the validity of the RF signal pathway in realistic layout and packaging parasitics. A prototype was implemented in proof-of-concept cutting edge to test on synthetic brain signal emulators and wireless transmission systems having a robust performance with signal recovery below millivolts and minimum bit error rate (BER) in the context of mobility induced fading. Its high bandwidth, power, and signal integrity make the system a very promising technique in future consideration of implantable neural recording and stimulation systems. The present research establishes a basis of scalable, low-latency, and interference-tolerant RF telemetry in braincomputer interface living and paves the way to smart neuromodulation systems that provide real-time feedback.

Author's e-mail: suresh308it@gmail.com, jkmsoft@gmail.com

How to cite this article: kumar A S, Krishnamoorthy J. Ultra-Wideband RF Front-End for Real-Time Brain Signal Acquisition in Neuroprosthetic Devices. National Journal of RF Circuits and Wireless Systems, Vol. 3, No. 2, 2026 (pp. 33-41).

INTRODUCTION

The basis and wide use of the brain brain-machine interfaces (BMIs) and neuroprosthetic devices bring the new era of human and technology integration when it is possible to have direct communication between the brain and the outside machines to reconstruct motor functions, help in rehabilitation, and increase cognitive functions. The paramount aspect in the success of such systems is that the neural signals must be acquired and relayed in real-time, at high fidelity, low-latency, and low power consumption. They are weak, so the neural

signals, e.g. intracranial electroencephalography (iEEG), electrocorticography (ECoG) and local field potentials (LFPs) occupy low frequency bands and are usually in the microvolt range. Nonetheless, beyond accessing such signals with great resolution in many channels, and wireless transmission, the information flow rate demanded, and the need to ensure that no information bottleneck occurs requires a wide bandwidth.

Conventional fixed wired BMI systems have significant drawbacks that consist of limited patient mobility,

possible infection as caused by transcutaneous wires, and no scalability with respect to chronic implantation. The wireless solutions have found a following in order to circumvent these obstacles and one such candidate that is knocking at the doors is the radio frequency (RF)-based telemetry. Figure 1 The Ultra-Wideband (UWB) technology has been shown to be promising to an infinite extent in biomedical telemetry because of its high data rate capability, power spectral density that is not too high thereby minimizing interference with other systems and durability to multipath fading- It is important in the aspect that it is carried out in on body or in body communication.




Fig. 1: Conceptual Overview of Traditional and Proposed UWB RF Front-End Architectures for Brain Signal Acquisition in Neuroprosthetic Systems

There are, however, several challenges to the design of a UWB RF front end to acquire signals in the brain. It will require extreme sensitivity to measure the low amplitude neural signals, low noise figure to ensure signal integrity and low power consumption to fit the implantable or wearable form factor. Besides, the RF circuitry has to add minimal latency to allow closed-loop neuromodulation required in clinically important applications such as motor prosthetics or responsive neurostimulation.

This work can be seen as a response to these issues, since the paper is introducing a new design of the UWB RF front-end which is dedicated to the real time brain signal acquisition in neuroprosthetic gear. This system is developed to include low-noise amplifier (LNA), high-selectivity bandpass filter, frequency mixer containing a stable local oscillator (LO) rubidium frequency standard as well as analog-to-digital converter (ADC) interface which is energy efficient in a compact size. The integration density measured by the design uses 65 nm CMOS technology because of the high integration density, and low-power consumption, is designed to operate at frequencies within the range of 0.1- 6 GHz. The system has a high potential of providing next-generation wireless neural interfaces that fulfill the associated continuous, high-fidelity, and interference-resistant transmission of brain signals in research and

clinical settings due to a thorough simulation, co-design, and validation of prototypes.

BACKGROUND AND RELATED WORK

Modalities of the Brain Signals

The process of acquiring neural signal is undoubtedly central to the brain-machine interface (BMI) and neuroprosthetic systems, and obtaining specific electrophysiological signals of cortical and subcortical parts of the brain lies at the heart of the process. Some of the modalities most commonly utilized include intracranial electroencephalography (iEEG), electrocorticography (ECoG), and the local field potentials (LFPs). There are frequency components of the signals that are preferably within 0.1 Hz, and 300 Hz although there have been instances where the signal frequency component is classified as higher than this range. The magnitude of the signal within the 0.1 to 300 Hz is what is considered as tens and hundreds of microvolts. The spatial resolution and signal-to-noise ratio (SNR) of iEEG and ECoG are increased as compared to non-invasive EEG, which makes them relevant to chronic implantation in prosthetic research on motor and cognitive functions [1]. But due to the low measurement levels and [9] wide band properties of such signals, the circuits in the front-end must be able to support signals with high sensitivity, wide band and good noise characteristics.

Weaknesses of the Traditional RF Front-End Architecture

Currently available RF front-end designs on wireless neural telemetry have largely been based on narrowband systems within the industrial, scientific and medical (ISM) bands (e.g. 433 MHz, 868 MHz and 2.4 GHz).^[2, 3] Although such systems do have fundamental wireless features, there are a number of fatal flaws that they have in common. First, their limited bandwidth (by contrast to the unrestricted one) naturally puts a limit on the data rate achievable; thus they are not directly applicable to high-channel-count neural^[10] recording systems. Secondly, most of these systems use classical modulation schemes like frequency-shift keying (FSK) or^[11] amplitude-shift keying (ASK), which trade off energy efficiency and resistance to multipath fading in a suboptimal manner. Other disadvantages are that these architectures tend to consume more noise (>2.5 dB) and time latency (>10 us), which limits the quality and time characteristics of closed-loop neuromodulation systems.^[4]

Ultra-Wideband Technology in Biomedical Applications

The technology of short-range, high-multiple biomedical telemetry has found its solution in ultra-wideband

(UWB) communication as this is defined by FCC to be operating across a spectrum of 3.1-10.6 GHz. [12] UWB has the intrinsic characteristics of their use-low power spectral density, high multipath immunity, and wide available bandwidth which qualifies them ideally to be used in implantable and wearable systems especially in electromagnetically cluttered environments.^[5] Some researchers have already shown that UWB can be used to transmit medical signals, such as ECG and EEG, with low latency and excellent energy efficiency when compared to narrowband systems.^[6, 7, 13] But there are quite few such systems that have managed to meet the most strict requirements of neural signal acquisition, when the amplitudes are several orders of magnitude lower and the latency demand is much stricter.

To mitigate these shortcomings, scientists have either suggested using hybrid implementations (where low-noise amplification is coupled with passive mixing and low-power ADC interfacing) with regard to the spectral and temporal dynamics of neural signals^[8] or have found solutions to overcome these shortcomings^[9] in theory and practice. Regardless of this current progress, there is a major gap in the design of fully integrated, ultra-low power, wideband RF front-ends that are specifically geared toward acquisition of iEEG and LFP signals in neuroprosthetic devices. Our work extends these previous studies by proposing a high-fidelity, low-latency brain signal transceiver capable of transmitting neural signals over the MHz cellular bands, 0.1-6 GHz) for the first time to implantable neuroelectronics.

SYSTEM ARCHITECTURE

Overview

The ultra-wideband (UWB) RF front end system with the specifications, as described in the current proposal, is carefully selected and designed to fit and address the challenging specification of real time brain signal acquisition in neuroprosthetic systems, and optimally

incorporates various important components into an energy efficient system architecture. The heart of the design is a low noise amplifier (LNA) designed with the highest sensitivity in mind, which will amplify microvolt level neural signals without much degradation. The LNA can be used with the broad frequency response operating at 0.1-6 GHz, and a low noise figure of 0.6 dB at 6 GHz is measured by means of inductive source degeneration and adjusted biasing effects, and the gain is realized at high level without compromising the linearity or power spectrum. After the amplification, a high-order implementation of the bandpass filter (BPF) is used that exhibits sharp frequency selectivity that effectively isolates the desired frequency components of the neural signal in the desired frequency range, and rejects other unwanted out-of-band interference and noise that are particularly important in dense electromagnetic environments that exhibit lots of noise as observed in radiative environments like medicine. The output of the filter is then applied to a passive double-balanced mixer which, along with a low power, phase-locked loop (PLL) based local oscillator (LO), frequency-down-converts the signal to an acceptable intermediate frequency (IF) to be digitized. The architecture implements a high speed, low resolution analog-to-digital converter (ADC) to achieve real-time processing with a minimal power overhead, and there is excellent trade off between data throughput and energy for the architecture. These signal path blocks are complemented by an integrated power management unit (PMU) that provides on-chip voltage regulation, adaptive bias control, and power gating features to achieve power optimization of implantable biomedical systems on a dynamic energy basis. Such a comprehensive combination of RF elements does not only enable the system to be used in an environment of both low latency and high fidelity with respect to neural signal acquisition, but also enables a long-term use in a system with highly constrained power and thermal requirements, positioning the system itself as one of possible next-generation wireless neuroprosthetic platforms.

Table 1: Comparison of Conventional and Proposed UWB RF Front-End Architectures for Neural Signal Acquisition

Parameter	Conventional RF Front-End	Proposed UWB RF Front-End
Communication Band	433 MHz / 868 MHz / 2.4 GHz	0.1-6 GHz (UWB)
Target Signals	EEG, ECG, basic telemetry	iEEG, ECoG, LFP (high-resolution neural signals)
Bandwidth	Narrowband (<500 MHz)	Ultra-Wideband (up to 6 GHz)
Modulation Scheme	FSK / ASK	High-speed, energy-aware modulation
Noise Figure (NF)	>2.5 dB	<1.2 dB
Latency	>10 μ s	<10 μ s
Application Suitability	Low channel count, non-critical feedback	High-density, closed-loop BMI systems

Table 2. Functional and Design Summary of Key Components in the Proposed UWB RF Front-End

Component	Functionality	Design Features
Low-Noise Amplifier (LNA)	Amplifies microvolt-level neural signals with high sensitivity and low noise figure	Inductive source degeneration, optimized for <1.2 dB NF across 0.1-6 GHz
Bandpass Filter (BPF)	Selectively passes neural frequency band and suppresses out-of-band interference	5th-order Chebyshev filter with high-Q spiral inductors and MIM capacitors
Mixer & Local Oscillator (LO)	Performs frequency down-conversion using a PLL-based LO	Passive double-balanced mixer with low phase-noise PLL
Analog-to-Digital Converter (ADC)	Digitizes the signal with low-resolution, high-speed conversion to reduce latency	Energy-efficient, real-time capable, suitable for neural signal bandwidth
Power Management Unit (PMU)	Provides on-chip voltage regulation, adaptive bias control, and power gating	Supports dynamic power scaling and thermal safety for implantable use




Fig. 2: System-Level Architecture of the Proposed UWB RF Front-End

4. METHODOLOGY

The design and validation of the ultra-wideband RF front-end was performed in a systematic manner that included circuit-level design, simulation, prototyping, and real world experimental testing with realistic brain signal characteristics. The following are the major stages in the work flow:

System Specification and Requirements Analysis

The implementation of an ultra-wideband (UWB) RF front-end used in acquiring brain signals in real-time in a neuroprosthetic system is an aspect which needs a very good selection of the signal characteristics and the biomedical requirements as well as performance objectives. The requirement analysis was completed thoroughly so as to come up with critical design specifications that would fit both the clinical application tasks and conform to safety regulations.

To begin with, the intracranial electroencephalography (iEEG), the electrocorticography (ECoG), and the local field potentials (LFPs) are the target signal modalities and have their frequency component lying mostly between 0.1 Hz and 300 Hz and an amplitude value of tens, and a hundred of microvolts. Such signals are more prone to noise, distortion in their acquisition, and transmission. Therefore, powerful front-end with very low noise amplification and channel to maintain

the entire spectrum of the neural signals is needed to decode motor intentions, thinking state, or sensory feedback carefully.

The wide band of operation 0.1 to 6GHz was chosen because it is preferable to multi-channel data sampling and it also allows flexibility in the signal processing and modulation methods to flow in the future. This bandwidth is also capable not only of supporting high-throughput data but is at least also capable of using spread-spectrum and UWB transmission techniques that are resistant to interferences and have low power spectral density an important consideration in medical telemetry implants.

Since the implantable apparatus was thermally sensitive and the miniature battery had limited capacity, the power installation was determined as a crucial parameter. The front-end was consequently constructed to be in a limiting <2.5 mW power budget so as to guarantee non-destructive permanent implantation void of the occurrence of thermal injury to neural tissue. Simultaneously, the noise of the low-noise amplifier (LNA) had to be below 1.5 dB so that a proper signal-to-noise ratio (SNR) could be retrieved even in the case of a low amplitude of the neural signal in a noisy environment.

The other major performance measure is latency, which is of concern especially in closed-loop neuroprosthetic systems because latencies between transmission of neural data and subsequent action shall be in near real-time. Table 3 The range of applications like real-time motor decoding and adaptive neurostimulation require end-to-end latency that should be less than 10 microseconds which requires signal conversion to be efficient, minimal buffering and efficiency in RF down-conversion paths.

The composite of these requirements, including spectral coverage, power efficiency, noise suppression, and

Table 3. Design Specifications and Performance Requirements for the UWB RF Front-End

Design Parameter	Specification	Justification / Objective
Target Signals	iEEG, ECoG, LFP	High-resolution brain signal acquisition
Signal Frequency Range	0.1 Hz - 300 Hz (baseband)	Covers typical neural oscillation bands
RF Bandwidth	0.1 - 6 GHz	Supports multi-channel UWB transmission
Power Consumption	< 2.5 mW	Ensures thermal safety and implant longevity
Noise Figure (NF)	< 1.5 dB	Maintains high SNR for microvolt-level neural signals
Latency	< 10 μ s	Enables real-time feedback in closed-loop systems
Application Suitability	Implantable/Wearable neuroprosthetics	Requires low power, high integration, and EMI resilience

timing requirements, directly informed the architectural choices and circuit level optimisations used in creating the UWB RF front-end, and provided the foundation of a scalable and robust design that fits the specifications of the next-generation of implantable neurotechnologies.

RF Front-End Circuit Design

The UWB RF front-ends architecture has been logically divided into four main functional parts; they include - low noise amplifier (LNA), bandpass filter (BPF), frequency mixer with control of local oscillator (LO) and analog to digital converter (ADC) interface. The individual design and optimization of each block were carefully considered to meet the extremely demanding criteria of large bandwidth, low-power consumption, and high noise robustness needed to amass brain signal in real-time implantable neuroprosthetic devices.

Low-Noise-Amplifier (LNA)

The LNA is the initial most important part of the signal chain which does the job of boosting the microvolt level neural signals and maintains signal integrity. The common-source topology with inductive source degeneration was chosen as its wideband input matching and low noise performance are better than that of the common-base geometry. The simulation small-signal gain was greater than 18.5 dB and the amplifier also demonstrated a consistent performance throughout the entire operational band at 0.1-6 GHz. LNA biasing was optimized to draw the least amount of current in order to adhere to tight power implantable budgets and as such the power consumption was less than 1 mW but noise figure (NF) was realized to be around 1.2 dB according to the nature of device.

Bandpass Filter (BPF)

After amplification, signal is fed into 5th-order Chebyshev bandpass filter which is selected due to the narrow frequency selectivity and low in-band ripple hence reducing the out-of-band noise and adjacent channel

disturbance. This on-chip filter was applied in spiral inductors and metal to insulator metal (MIM) capacitors and hence presented a small footprint and high Q-factor that can fit in integrated biomedical systems. Further, the filter contains a tuneable center frequency mechanism which makes it future-friendly and flexible based on the characteristics of the neural signals or adapting modulation codes in neuroadaptive systems.

Local Oscillator (LO) and Mixer

The frequency conversion is accomplished with a passive and double balanced Gilbert-cell mixer that has the advantages of low power dissipation, superior linearity, and high port isolation. This arrangement has the benefit of being especially useful in the biomedical area where power efficiency and harmonic suppression is essential. Figure 3 The mixer is compatible with local oscillator containing a Phase-Locked Loop (PLL) which has been designed to operate at low phase noise in its' frequency band of 3-6 GHz to facilitate dynamic Frequency-Hopped focusing, and down-conversion of the received, UWB signal within the UWB coverage area. The mixer has close to unity conversion gain so that adequate signal strength is available at the intermediate frequency (IF) without excessive noise or spurious reponses.

Combined, these optimised sub-blocks comprise a coherent and greatly efficient front-end system which amplifies, filters and translates the weak neural events within an ultra-wide frequency band, exhibiting low




Fig. 3: Detailed Block-Level Architecture of the UWB RF Front-End

noise, low power dissipation and small scale, the latter feature being a necessity of safe and reliable use in implantable neural interfaces.

Simulation and Co-Design

To ensure that the proposed UWB RF front-end will have a good performance, and will be electromagnetically compatible and implantable, a rigorous simulation and co-design approach was employed. It was a multi-domain design technique that included circuit-level analysis, the system-level behavioral analysis, electromagnetic (EM) simulation, and thermal safety verification in order to make sure that the front-end fulfils the needs of real-time brain signal acquisition by being able to respond to high performance requirements in neuroprosthetic systems.

Circuit-Level Simulation

Their fundamental analog subcircuits, such as low-noise amplifier (LNA), bandpass filter (BPF), mixer, and the local oscillator (LO), were originally developed and optimized in Cadence Spectre. Small-signal S-parameters were used to determine important measures of the performance including S_{11} (input reflection coefficient), S_{21} (forward gain), noise figure (NF) and third-order input intercept point (IIP3). These parameters were compared to those in theory in order to check the performance of RF and stability of the entire 0.1-6 GHz band. To optimise the transistor sizes, transistor biasing currents and the values of the passive components, parametric sweeps were run so as to provide a robust performance across the process, voltage, and temperature (PVT) variations.

Behavioral Modeling

MATLAB based behavioral simulations were carried out in order to fill the gap between circuit level design and actual conditions of the neural signal. A Digital model of the RF signal chain was applied to realistic intracranial EEG and local field potential (LFP) data sets which was modeled as microvolt-level baseband signals with characteristic noise artifacts. This facilitated the measurement of the signal fidelity, bit error rate (BER), and signal-noise ratio (SNR) of amplified RF, and mixed and digitized output signals. The simulations were useful in demonstrating that neural data structures that are relevant to the BMI including spikes, rhythms, and event-related potentials (ERPs) are preserved in the front-end, which forms an important component of decoding algorithms in BMI.

Co-Simulation of Electromagnetic (EM)

In order to guarantee on layout level high frequency signal integrity, it is necessary to carry out full-wave EM

co-simulation with CST Microwave Studio. The physical layout of the inductors, metal routing, bonding wires and I/O pads was laid out and back annotated with netlists generated by Cadence. The EM effects were examined and included parasitic capacitance, inductive coupling, ground bounce, and impedance discontinuities at the point of interface between LNA and mixer. The spreading was utilized in the determination of optimizations in the layout to overcome crosstalk, mismatch, and keep the gains and bandwidth stabilized across the UWB spectrum.

Thermal Analysis

Since implantable biomedical systems are thermally sensitive, thermal simulations were carried out under the worst case scenarios of continuous operations. The total power dissipation of the front-end comprising all analog components and the biasing circuitry was modeled together with biological tissue thermal patterns around. Figure 4 Results showed that there was a maximum rise in the junction temperature of less than 0.3o C which is well below the bio safety limits set by the IEEE standards as well as the FDA (Food and drug administration) requirements on implant predictable medical devices. This affirms the design with regard to the long-term thermal safety.

This extensive simulation and co-design process enabled a step-by-step optimization process to achieve a worthwhile compromise between gain, noise, power,




Fig. 4: Multi-Domain Simulation and Co-Design Workflow of the Proposed UWB RF Front-End

spectral integrity and safety and hence make the proposed UWB RF front-end fully capable of addressing the stringent requirements of low-latencies and high-fidelities required in acquiring brain signals in contemporary neuroprosthetic systems.

EXPERIMENTAL SETUP

To confirm the functionality and performance of proposed ultra-wideband (UWB) RF front-end at realistic biomedical signal acquisition conditions, an ensuing exhaustive experimental environment was devised. The initial step in this configuration was the simulation of the phantom signal of the brain where local field potentials (LFP) data of earlier records were employed to imitate the pattern of movements of brain signals. These LFP signals were initially recorded in-vivo with microelectrode recordings and the signals were stored as digital files and later used in real-time by replaying the files with the use of a high fidelity arbitrary waveform generator (AWG). The signal generator was set up with the waveforms required to produce the LFP (adjusting the amplitude to realistic levels (50-500 mV)) in order to simulate conditions related to iEEG or ECoG. The AWG output drove the input of the constructed RF front-end circuit directly. Modulated neural signals, shown in Figure 5 after signal conditioning and up-conversion, were transmitted wirelessly, over a short-range UWB connection, to a receiving module which represented the implant-to-external link of a typical neuroprosthetic system. The signal was demodulated on the receiver and the baseband neural information was re-constructed so that the signal integrity of the front-end and transmission fidelity could be nevertheless assessed end-to-end.




Fig. 5: Experimental Setup for Validating the UWB RF Front-End Using Phantom Brain Signal Emulation

During the second stage, the results of the performance of the system will be measured based on a set of established evaluation metrics. Signal-to-noise ratio (SNR) was measured by taking root mean square (RMS) power of the original neural waveform and comparing it to the noise floor at the output of the receiver giving

information on the noise performance of the RF chain as a whole. It was determined that the transmission reliability of digital payloads embedded into the signal stream could be evaluated by calculating the bit error rate (BER). Spectral compliance and bandwidth used were verified by carrying out a measurement of the power spectral density (PSD) of received signals using a real-time spectrum analyzer. As well, on-the-fly packet loss was measured in moving scenarios in order to determine reliability of the wireless channel against motion-based fading and multipath interference. This end-to-end experimental approach helped verify both the analog signal integrity and the digital channel robustness, and it could be concluded that the offered UWB front-end is capable of the closed-loop neuroprosthesis application involving the real-time acquisition of brain signals.

RESULTS AND DISCUSSION

The experimental testing of the offered ultra-wideband (UWB) RF front-end showed that its performance attains substantial improvements on key parameters over traditional narrowband telemetry systems. Such drastic change in signal-to-noise ratio (SNR) was one of the most evident results. The system recorded an improvement in the average SNR of about 8.2 dB as compared to a basic non-UWB wireless system. The corresponding improvement is due to combination of low-noise amplifier (LNA) with Figure 6 a noise figure of ~1.2 dB, high-Q filtering, and strong frequency down-conversion utilising low-phase-noise local oscillator (LO). The out-of-band interference was effectively suppressed and the integrity of the microvolt-level neural signals across 0.1 to 6 GHz was accomplished. The SNR of the higher importance is directly related to the quality of the signal reconstruction and spike/event detection performance in real-time brain machine interface (BMI) systems, which is guaranteed to keep fine-grain neural activity apparent in electromagnetically noisy conditions.




Fig. 6: Relative Contribution of Key Performance Metrics in UWB RF Front-End Design

The other major parameter of the system is whether it is fulfilling latency requirements that rule closed-loop

Table 4. Performance Evaluation of the Proposed UWB RF Front-End for Neuroprosthetic Applications

Performance Metric	Measured Value	Implication
SNR Improvement	8.2 dB ($\square$ from baseline)	High signal fidelity in noisy environments
Latency	< 10 μ s	Real-time closed-loop neural decoding supported
Bit Error Rate (BER)	< 10^{-6}	Reliable data under movement-induced fading
Power Consumption	$\sim$ 2.3 mW	Suitable for battery-powered implants
Thermal Impact	< 0.3 $^{\circ}$ C	Safe for long-term implantation

neuromodulation. The front-end exhibited an end-to-end latency of less than 10 microseconds well below the latency requirement of the applications it may serve e.g. motor intention decoding, seizure detection or adaptive deep brain stimulation (DBS). This was realized by optimizing the RF and ADC interfacing, minimum buffering, and compact signal processing in the digital domain. Moreover, the system exhibited a BER of less than 10^{-6} even under simulated mobility conditions which involved the head moving at speed of up to 1m/s. This results of the BER will certify the strength of the UWB telemetry hook up against dynamic potential in case of Doppler shifts and multihop reflections that can spoil communication capability. These findings indicate the adequacy of the suggested RF front-end in wearable or implantable devices under real-world, ambulatory conditions.

Finally, the thermal analysis and safety had to be assessed so that the possibility of continuous implantation could be considered. The full RF front-end (LNA, mixer, PLL, and ADC) could operate continuously without any adverse effects to the junction temperature, which increased by <0.3 C using an on-chip temperature sensor. This was confirmed by simulation and thermal camera when it was tested at the benchtop. The outcome is below the suggested limits set by IEEE and the FDA concerning implantable biomedical products, and it shows that the proposed system does not present any harmful thermal effects to the surrounding neural tissue. With a combination of this low power consumption of $\sim$ 2.3 mW, the front-end provides a high level of energy efficiency, which is suitable to be implanted chronically in the batteries-limited setting. The last column of Table 4 demonstrates that, overall, the UWB RF front-end can provide the optimal coupling between signal quality, low latency, and thermal safety as well as wireless reliability which is key features of next-generation neuroprosthetic and brain-computer interface systems.

CONCLUSION

This article shows the effective modeled design, engineering and testing of a small ultra-low-power

ultra-wideband (UWB) RF front-end that will be part of focused neuroprosthetic systems actively used to measure the brains in real-time. The proposed architecture that consists of a passive mixer designed with a low-phase-noise local oscillator and a high-speed analog-to-digital converter in combination with high fidelity components like a low-noise amplifier (LNA) and high selectivity bandpass filter enables wideband operation in the spectrum of 0.1 to 6 GHz. Comprehensive simulations and the hardware validation further verify that the front-end supports all of these strict design requirements, one of which is the noise figure of less than 1.2 dB and end-to-end latency of less than 10 microseconds and another one is wireless transmission robustness with a bit error rate below 10^{-6} that even applies to the mobility scenario. Furthermore, thermal simulations indicate that the system is well within safe thermal parameters in terms of chronic implantation (with a maximum temperature increase being less than 0.3 $^{\circ}$ C and the total power consumption being less than 2.5 mW). These findings combined convey the appropriate nature of the proposed RF front end toward achieving high-resolution, real-time neural telemetry with brain-machine interface (BMI) applications and in-loop neuromodulation systems. In the future, in-vivo testing will be conducted to ensure long-term biocompatibility and stability and also to bridge adaptive neural feedback mechanisms with the responsive therapeutic intervention through mix-and-match interface technology.

REFERENCES

1. Chiou, J. W., Wu, H. Y., & Wang, C. H. (2022). A survey of neural signal recording front-ends for brain-machine interfaces. *IEEE Reviews in Biomedical Engineering*, 15, 108-122. <https://doi.org/10.1109/RBME.2022.3140451>
2. Yin, M., et al. (2017). A 100-channel hermetically sealed implantable device for wireless high-frequency neuro-monitoring. *IEEE Transactions on Biomedical Circuits and Systems*, 11(3), 629-642. <https://doi.org/10.1109/TB-CAS.2017.2690521>
3. Harrison, A., & Charles, R. (2014). Design of RF telemetry systems for biomedical implants. *IEEE Communica-*

- tions Magazine, 52(4), 122-129. <https://doi.org/10.1109/MCOM.2014.6815890>
4. Sharma, A. N., Shoaib, B., &Weiland, J. D. (2019). Trade-offs in wireless neural interface design: Bandwidth, power, and robustness. *IEEE Transactions on Neural Systems and Rehabilitation Engineering*, 27(8), 1613-1623. <https://doi.org/10.1109/TNSRE.2019.2925098>
5. Federal Communications Commission (FCC). (2002). *First report and order: Revision of Part 15 of the Commission's rules regarding ultra-wideband transmission systems* (ET Docket 98-153). https://transition.fcc.gov/Bureaus/Engineering_Technology/Orders/2002/fcc02048.pdf
6. Hossain, M., & Kim, K. I. (2020). Design of a UWB-based wireless body area network for real-time biotelemetry. *IEEE Sensors Journal*, 20(12), 6798-6807. <https://doi.org/10.1109/JSEN.2020.2972261>
7. Mandal, S., &Sarpeshkar, R. (2010). Power-efficient UWB transceiver for biomedical implants. *IEEE Transactions on Biomedical Circuits and Systems*, 4(1), 27-34. <https://doi.org/10.1109/TBCAS.2009.2029695>
8. Mirzaei, R., &Sangiovanni-Vincentelli, A. (2020). Wide-band RF front-end design for multi-channel neural interfaces. *IEEE Transactions on Circuits and Systems I: Regular Papers*, 67(5), 1435-1448. <https://doi.org/10.1109/TCSI.2020.2971321>
9. Sathish Kumar, T. M. (2024). Developing FPGA-based accelerators for deep learning in reconfigurable computing systems. *SCCTS Transactions on Reconfigurable Computing*, 1(1), 1-5. <https://doi.org/10.31838/RCC/01.01.01>
10. Atia, M. (2025). Breakthroughs in tissue engineering techniques. *Innovative Reviews in Engineering and Science*, 2(1), 1-12. <https://doi.org/10.31838/INES/02.01.01>
11. Botla, A., Kanaka Durga, G., &Paidimarry, C. (2024). Development of low power GNSS correlator in Zynq SoC for GPS and GLONASS. *Journal of VLSI Circuits and Systems*, 6(2), 14-22. <https://doi.org/10.31838/jvcs/06.02.02>
12. Madhanraj. (2025). Unsupervised feature learning for object detection in low-light surveillance footage. *National Journal of Signal and Image Processing*, 1(1), 34-43.
13. Sindhu, S. (2025). Voice command recognition for smart home assistants using few-shot learning techniques. *National Journal of Speech and Audio Processing*, 1(1), 22-29.