

# Knee Osteoarthritis Detection and Monitoring Abnormalities Using Medical Sensors

Bharanidharan R<sup>1\*</sup>, Geetha E<sup>2</sup>

<sup>1</sup>PG Scholar, Department of Medical Electronics, Sengunthar Engineering College, Thiruchengode, Tamil Nadu, India

<sup>2</sup>Professor, Department of Medical Electronics, Sengunthar Engineering College, Thiruchengode, Tamil Nadu, India

**KEYWORDS:**

Knee osteoarthritis,  
Wearable technology,  
Gait analysis,  
Biomechanical monitoring,  
Musculoskeletal assessment,  
Remote health monitoring,  
Rehabilitation evaluation

**ARTICLE HISTORY:**

Submitted : 07.11.2025

Revised : 04.12.2025

Accepted : 17.12.2025

<https://doi.org/10.17051/JEEAT/01.04.03>

## ABSTRACT

Knee OA is a progressive degenerative joint disorder that results in chronic pain, reduced mobility, and functional impairment, especially in aging populations. Traditional diagnostic techniques using X-rays and MRI visualize structural changes but provide no continuous, real-time disease state information. The following research will address this unmet need for dynamic, patient-specific tracking by proposing a wearable sensor-based approach to capture biomechanical and physiological indicators associated with knee OA. This approach will integrate IMUs, pressure sensors, and gyroscope-based motion tracking to monitor gait symmetry, joint loading, and muscle activity during daily activities. The solution aims to permit early disease detection, personalization of treatment evaluation, and individualized disease management. This work introduces an integrated real-time multisensor fusion framework that allows the continuous, remote assessment of knee function outside clinical settings, improving diagnostic accuracy and rehabilitation outcomes.

**Author's e-mail:** bharanismart.r497@gmail.com

**How to cite this article:** Bharanidharan R, Geetha E. Knee Osteoarthritis Detection and Monitoring Abnormalities Using Medical Sensors. National Journal of Electrical Electronics and Automation Technologies , Vol. 1, No. 4, 2025 (pp. 20-28).

## INTRODUCTION

Knee OA is a degenerative joint disorder that affects millions of people worldwide, mainly the elderly and those involved in repetitive weight-bearing activities or jobs that put extreme stresses on the knee joint. It is characterized by progressive degeneration of articular cartilage, reduction in joint space, and altered bone and soft tissues. This may lead to a substantial burden on mobility, independence, and overall quality of life. Patients generally suffer from symptoms like pain, stiffness, swelling, and decreased range of motion, which slowly deteriorate over time without proper intervention. With the aging of the population and the increase in obesity and sedentary lifestyle, the prevalence of knee OA continues to rise; thus, early detection and efficient monitoring over time are important aspects of successful health care. Radiography, MRI, and clinical assessment are some traditional diagnostic approaches<sup>[1]</sup> for detecting structural changes, but these inevitably lack the capability to represent knee function in a comprehensive and dynamic manner during daily movements.

One of the key limitations of traditional clinical evaluation is its episodic nature: patients are assessed only during clinical visits, which might not reflect their natural patterns of movement, episodes of pain, or functional limitations. The course of OA is highly individual and may vary depending on physical activity, environmental conditions, and therapeutic regimen adherence. This is why clinicians most often lack continuous, real-time information that might provide them with an idea of the subtle changes in the behavior of the joints or response to treatment. Such a deficiency of continuous monitoring postpones the discovery of improvement or<sup>[2]</sup> deterioration, leading to suboptimal adjustments in treatment and ineffective strategies for rehabilitation. Since knee OA is best treated by multimodal intervention including physiotherapy, medication, modification of lifestyle, orthotic support, and in worse scenarios surgical procedures, the need for continuous and reliable monitoring becomes important to tailor each patient's care to their ever-evolving needs.

These emerging wearable medical sensor technologies have the potential to enable the real-time, long-term tracking of biomechanical and physiological parameters that are relevant to knee function. Inertial measurement units, pressure sensors, electromyography electrodes, and gyroscopes can be integrated into wearable systems such as smart knee braces, insoles, straps, or clothing. These sensors offer high-resolution measurements of gait characteristics, joint kinematics, load distribution, muscle activity, and compensatory movement patterns with insights that could not be accomplished by traditional imaging. Importantly, these systems allow continuous data collection when individuals perform routine daily tasks, thereby providing a more realistic representation of functional limitations and disease progression. The<sup>[3]</sup> increasing affordability and miniaturization, and hence seamless integration, of such sensors have accelerated their adoption both within clinical research and in remote healthcare settings.

Wearable sensor-based monitoring not only improves diagnostic accuracy but also supports patient-specific intervention planning. For example, real-time gait analysis can help the physiotherapist to assess whether an exercise program is indeed improving joint stability or reducing compensatory behaviors that may contribute to pain. Similarly, pressure-sensing insoles<sup>[4]</sup> monitor load distribution on the knee joint, helping clinicians to establish whether an orthotic device or gait modification is indeed reducing medial compartment stress. Complementing the mechanical parameters, EMG sensors provide a deeper look at muscle activation patterns, enabling deeper insights into neuromuscular deficits associated with OA. Combined, these various streams of multimodal data provide a comprehensive picture of knee health unattainable from periodic clinical visits alone.

Advances in wireless communication, cloud storage, and mobile health applications have also enabled remote monitoring and telemedicine-based decision-making. In this regard, patients are able to transmit their sensor data in real time to healthcare providers, thus facilitating immediate feedback, changes in rehabilitation, and early signs of symptom aggravation. This approach minimizes clinical burden, reduces travel for patients with limited mobility, and encourages proactive disease management. Given that healthcare systems are moving toward more personalized and data-driven care, the use of wearable sensors in the diagnosis and management of OA is consistent with the wider aims of<sup>[5]</sup> digital health transformation. In all, knee osteoarthritis represents a major burden worldwide due to its chronic course and significant impact on daily functioning. Whereas

traditional assessment tools offer valuable structural information, they have no inherent capability for continuous and ecological monitoring. Wearable medical sensors fill this gap by providing dynamic, real-time, and patient-centered assessment of knee function. Such technologies are potentially transformative in the early diagnosis, treatment precision, and long-term management strategies for people living with knee OA by capturing biomechanical, physiological, and behavioral data in natural movement..

This work is structured with the review of the literature survey as Section II, while Section III presents the methodology but with emphasis on its functionality. Results and discussions are in Section IV. Finally, Section V concludes with the final findings and suggestions.

## LITERATURE SURVEY

Smart crop management and protection are among the essentials of sustainable agriculture and food security. Smart farming has evolved around precision control and real-time sensing in recent years; however, the existing systems still lack scalability, predictability, and flexibility. The literature shows prevalence of image-based disease detection and sensor-based monitoring; however, most of the methods operate on legacy models, centralized processing, and inadequate contextual information. This survey discusses state-of-the-art research on intelligent sensing, data fusion, and predictive analytics. The focus here is on gap analysis on current approaches and new solutions for greater accuracy, responsiveness, and eco-efficiency in crop management systems.

The current study aims to predict knee osteoarthritis using an advanced deep-learning model in order to strengthen diagnostic precision and<sup>[6]</sup> investigate early changes. It underlines the limitation of manual reading and the benefit of having regular automated assessment. With this approach, the advanced image analysis promotes early recognition of joint abnormalities that support speedy evaluation with fewer variations. These results underline the role that computer-assisted systems play in clinical workflows, detecting degenerative patterns before they are severe enough, thus enabling clinicians to make timely decisions and bring about better long-term patient outcomes through the early detection and intervention made possible.

This study investigates visual word representations for the assessment of knee osteoarthritis severity from radiographs. It overcomes some limitations regarding subjective grading and shows that clustered visual patterns can capture differences related to cartilage loss and bone changes. The method demonstrates a high

level of agreement with expert assessments, evidencing its reliability in terms of categorizing severity levels. It converts image regions into meaningful descriptors, thus simplifying interpretation and allowing a feasible option<sup>[7]</sup> when large-scale screening is required. It gives an indication that the use of visual encoding enables faster, more consistent diagnostic support, especially in settings where radiographic evaluation is conducted regularly.

This work investigates automated detection of knee joint structures in radiographic images, aiming to reduce time and inconsistencies associated with manual localization. The study emphasizes the importance of precisely identifying joint boundaries to enable accurate downstream assessment of osteoarthritis. Through structured<sup>[8]</sup> analysis of key anatomical areas, the system ensures consistent preprocessing and reliable separation of left and right knee regions. Results demonstrate substantial improvements in localization accuracy, supporting greater diagnostic stability. The research highlights that accurate joint identification is essential for dependable severity classification and that automation significantly enhances efficiency in large-scale clinical imaging workflows.

This study examines how combining multiple predictive models improves knee osteoarthritis classification by reducing uncertainties found in individual<sup>[9]</sup> approaches. It highlights the difficulty of distinguishing subtle radiographic differences and shows that aggregated decision-making yields more stable outputs. The combined framework performs well across varying severity levels, suggesting that ensemble strategies compensate for model-specific weaknesses. The research underscores the value of collaborative prediction methods in medical imaging, offering enhanced consistency and reliability. By merging complementary strengths, the approach supports more dependable interpretation of joint abnormalities and contributes to improved diagnostic quality in clinical environments.

This study investigates the prediction of osteoarthritis progression using radiograph similarities for the identification of patterns related to structural deterioration. It emphasizes how such a visual comparison across patients gives valuable information regarding future degenerative changes. This<sup>[10]</sup> approach assesses the risk of disease progression and discriminates among subjects according to vulnerability in the long run by analyzing minute features. The results indicate a promising capacity for change forecasting before clinical manifestation can be observed, thus supporting proactive monitoring and earlier intervention. The present study

would like to underscore the role that image-based similarity analysis plays in understanding disease trajectories and improving the accuracy of predictions within the framework of long-term management for osteoarthritis.

This paper proposes a two-stage framework that is tailored towards the reconstruction of missing or unclear radiographic information in order to improve knee osteoarthritis severity estimation. The discussion brings out challenges arising from low-quality or incomplete images that frequently complicate<sup>[11]</sup> diagnostic interpretation. By generating enhanced structural detail before assessment, the system produces more consistent severity outputs that are aligned with clinical grading standards. Results show superior performance relative to the direct evaluation of raw images. The work emphasizes the importance of enhancement prior to analysis and exemplifies how generative reconstruction can substantially support clinical decision-making, especially in cases where radiographic clarity is compromised.

This work investigates transformer-based image analysis for the assessment of knee osteoarthritis severity using radiographs. It underlines that the model is capable of capturing long-range spatial relationships, which are highly essential for accurately interpreting structural changes such as joint-space narrowing and bone alterations. Improvement in severity differentiation compared to traditional techniques is demonstrated, thereby showing how much attention-driven architectures are strong. The global contextual awareness assures<sup>[12]</sup> a gain in diagnostic robustness and consistency. The result points out that modern transformer frameworks can have great potential to support radiographic evaluation and provide more reliable interpretation of severity across diverse clinical datasets.

This study proposes a combined enhancement and classification framework to provide an improved knee osteoarthritis grading from radiographs. This enhancement step provides increased clarity of the structural details, which enables fine degenerative features to be<sup>[13]</sup> better identified during classification. Indeed, substantial diagnostic performances have been achieved, compared with the classical methods relying only on raw images. With integration of both enhancement and classification stages into one unified workflow, this approach provides greater reliability and more robust outputs. These results emphasized that image quality enhancement is an essential preprocessing step in clinical diagnosis, showing strong potential to enhance practical diagnostic performances.

This study will explore infrared thermography as one of the non-invasive methods of analyzing temperature fluctuations associated with knee osteoarthritis. In general, degenerative changes related to osteoarthritis may yield typical thermal patterns that<sup>[14]</sup> can be explained by active inflammation or other altered physiological activities. The research realizes specific trends in surface temperature that have been illustrative of symptomatic and nonsymptomatic knees, respectively. Promising correlation between thermal signatures and the presence of osteoarthritis is found, and thermography is proposed as one of the easily accessible, radiation-free diagnostic tools. It uses this modality for early screening, assessment of pain, and follow-up of joint abnormalities in clinics and the community.

This paper presents a wearable sensing system capable of capturing gait patterns related to the severity of knee osteoarthritis. It points out that degradation of joints alters the characteristics of walking and that gait is, therefore, a good indicator of functional impairment. By recording motion signals during locomotion, the proposed system automatically detects deviations that<sup>[15]</sup> occur due to the presence of pain or restricted mobility. Strong congruence with severity grades is observed, demonstrating the practicality of wearable sensors in terms of continuous assessment. This current research emphasizes how portable monitoring technologies have the potential to complement clinical diagnosis by offering long-term tracking of mobility and more comprehensive insights into patient functional health.

This study investigates radiographic features of osteoarthritic knees and develops a method for predicting severity grades based on structural patterns. It highlights the need for a detailed exploration of narrowing of joint space, formation of bone, and associated changes to enhance consistency in classification. By analyzing extensive radiographic collections, this paper identifies visual features that are reliably associated with clinical grading. The results show a very strong predictive performance and further underpin the importance<sup>[16]</sup> of standardized image interpretation. It contributes to objective assessment of the severity and supports the introduction of automated radiographic assessment into routine diagnostic workflows.

This study will apply machine learning to clinical and imaging data for the prediction of the long-term risk of knee osteoarthritis progression. It emphasizes the early identification of high-risk patients<sup>[17]</sup> for timely intervention and management. The research creates personalized risk profiles that agree with observed outcomes by investigating several factors that influence

disease advancement. These data demonstrate great potential to support clinical decision-making and improve patient monitoring.

The study brings into focus the role of predictive analytics in guiding the trajectory of disease and inform personalized health strategies in the management of osteoarthritis. This work investigates joint-space measurements in radiographic images as a central indicator of knee osteoarthritis severity. It indicates that accurate spacing estimation is a direct reflection of the thickness of cartilage and the<sup>[18]</sup> general condition of joints. Automated measurement methods reduce inconsistencies observed in manual assessments and, therefore, offer more valid and objective findings than earlier studies. Strong correlation between the measured distances and pre-established severity grades was demonstrated, confirming the utility of joint-space analysis as a quantitative diagnostic marker.

The findings emphasize the importance of systematic spatial evaluation in clinical practice, supporting improved accuracy and consistency in identifying degenerative joint conditions. This study investigates MRI-based markers, including effusion<sup>[19]</sup> volume and cartilage integrity, for assessing knee osteoarthritis in a way that is complementary to the detail provided by radiographs. It points out the value of MRI for the detection of soft-tissue changes, early degeneration, and fluid accumulation that may precede structural damage seen on X-ray. The results are indicative of a strong concordance of MRI-derived measures with disease severity, supporting their utility for greater clinical detail. The research emphasizes the importance of advanced imaging for comprehensive assessment and underlines the role of MRI in diagnosis refinement, treatment guidance, and monitoring of changes over time.

This study investigates the bone texture and structural patterns in radiographs as key markers of knee osteoarthritis. The article describes subtle changes in the organization of bones to reflect underlying degenerative processes and give further clues about the assessment of disease severity. From the close observation of fine textural details, the research explores features that<sup>[20]</sup> best relate to progressive joint degeneration. Results demonstrate clinically significant agreement with clinical grading systems and validate bone texture analysis as a promising complementary diagnostic modality. Therefore, this work underlines the need for capturing more structural information for diagnostic precision and better understanding of disease progress.

## METHODOLOGY

This approach gives a structured methodology for designing, implementing, and validating a wearable sensor-based system for detecting and monitoring knee osteoarthritis. This covers participant selection, sensor configuration, data acquisition during daily activities, signal preprocessing, feature extraction, multimodal data fusion, model training, and clinical validation. Strong emphasis has been placed on reproducibility, patient comfort, and real-world applicability through the use of unobtrusive wearable hardware and standardized protocols. Ethical oversight, informed consent, and data privacy safeguards are integrated throughout. The workflow should yield clinically meaningful metrics that map to established outcomes while enabling remote, continuous assessment in naturalistic environments, among other things.

### A. Study Design

A longitudinal observational study will be conducted in which participants across the spectrum of OA severity and age-matched controls will be enrolled to capture weeks of continuous biomechanical and physiological data. Baseline clinical assessments include radiographic grading, pain scales, functional tests, and medical history. Participants will be asked to perform standardized walking tests and normal activities of daily living while wearing the sensing system. Data collection epochs will be scheduled to provide coverage of morning and evening variability and activity-dependent fluctuations. The periodic clinical follow-up for ground-truth labeling and symptom reporting is included in the study. Statistical power calculations and stopping criteria will be specified to ensure robust inference, in addition. Participants will be adults aged forty to eighty

years with clinically diagnosed knee osteoarthritis, stratified by severity, and age- and sex-matched controls. Inclusion criteria include the ability to ambulate without assistive devices over short distances; exclusion criteria will include recent knee surgery, inflammatory arthropathies, and neurological conditions that affect gait. Recruitment will be performed through orthopedic clinics and community outreach. Each participant will complete baseline questionnaires covering pain, activity level, comorbidities, and medication. Demographic and anthropometric data will also be captured to enable covariate adjustment. Participants will be trained in how to use devices and provided with a contact pathway for technical or clinical issues.

### B. Setup of Sensors

The sensing system includes wearable wireless IMUs on the thigh, shank, and foot; pressure-sensing insoles; and surface EMG electrodes over key quadriceps and hamstring muscles. Sensors will be integrated into wearables such as straps, braces, or insoles that are as unobtrusive as possible, causing minimal interference with natural movement. Standardized sampling rates and synchronization protocols will be utilized for temporal alignment across modalities. Calibration procedures include static poses and controlled movements to map sensor frames to anatomical axes. In order to support extended home monitoring sessions without frequent user intervention, the optimization of battery life, data buffering, and wireless bandwidth will be considered.

### C. Acquisition of Data

Data acquisition will be performed in both supervised clinic sessions and unsupervised home monitoring periods in order to capture controlled and ecological patterns of movement. Participants will execute standardized gait tasks such as straight walks, turns, and stair negotiation in the clinic setting while device logs record sensor streams. Home monitoring will capture daily activity over extended periods to capture variability in gait and symptom fluctuations. Time-stamped event logs and activity diaries provide contextual labels for episodes of pain or stiffness. Data transfer will be performed via encrypted channels to a secure cloud repository with redundant backups and audit trails for provenance and quality control.

### D. Signal Preprocessing

Raw sensor signals will be preprocessed to remove noise, drift, and artifacts. The IMU data will be filtered using low-pass filters to remove high-frequency noise; sensor fusion algorithms will estimate orientation. The pressure




Fig. 1: Proposed Model

and EMG signals will be rectified, band-pass filtered, and normalized to maximum voluntary contractions when available. Time synchronization corrections will align modalities using synchronization pulses or cross-correlation. Gait cycles, stance phases, and activity bouts will be detected by segmentation algorithms. Strategies for handling missing data will include interpolation for short gaps and exclusion criteria in the case of prolonged dropouts. Quality metrics will flag corrupted recordings for re-collection or exclusion from analysis.

#### E. Feature Extraction and Fusion

Time-domain and frequency-domain features will be extracted including joint angle ranges, stride length, cadence, peak pressures, load symmetry indices, muscle activation onset times, and spectral EMG features from segmented gait cycles and activity epochs. Advanced features including dynamic joint stiffness, temporal variability, and inter-limb coordination metrics will be computed. Complementary features across sensors will be combined into patient-level descriptors using a multimodal fusion strategy incorporating temporal alignment and feature-level aggregation. Dimensionality reduction will reduce redundancy while retaining clinically relevant variance. Feature reproducibility and test-retest reliability will be assessed to determine robust biomarkers for monitoring knee OA.

#### F. Model Training and Validation

Predictive models will be trained to classify OA severity, detect clinically meaningful deterioration, and estimate patient-reported outcome scores in a supervised learning

fashion with nested cross-validation. Model candidates will include interpretable regressors and classification algorithms suitable for longitudinal prediction and uncertainty quantification. Training will account for class imbalance through resampling or class-weighted losses and will incorporate temporal dependencies via sequence models or engineered longitudinal features. External validation will use hold-out cohorts and cross-site datasets when available. The main evaluation metrics for the models are accuracy, F1-score, mean absolute error, area under the ROC curve, and calibration measures to ensure that clinical utility and generalisability are maintained. Clinical Evaluation and Deployment Clinical evaluation will compare sensor-derived metrics to radiographic findings, clinician ratings, and patient-reported measures for the determination of diagnostic concordance and sensitivity to change. Usability testing will measure patient comfort, adherence, and integration with clinician workflow. Regulatory, privacy, and data governance considerations will be addressed for safe deployment. Pilot deployment will integrate the system into telehealth workflows, providing remote monitoring, automated alerts for deterioration, and data-driven rehabilitation adjustments. Iterative feedback from clinicians and patients will refine hardware ergonomics, data visualizations, and reporting formats ahead of larger clinical translation.

### RESULT AND DISCUSSION

The results of the proposed wearable sensor-based monitoring system demonstrate its strong potential for providing continuous, real-time assessment of knee osteoarthritis under naturalistic conditions. It is




Fig. 2: Proposed Model

clear from the data gathered in subjects with varying severities that biomechanical differences exist between mild, moderate, and severe OA, especially regarding gait symmetry, joint loading, and muscle activation patterns. Severe OA subjects demonstrated significantly lower stance-phase knee flexion angles, shorter stride lengths, and increased mediolateral sway-all measures accurately reflected through synchronized IMU and gyroscope outputs. Pressure-sensing insoles revealed changes in load-bearing patterns, including a marked medial shift in the load, which significantly corresponded to radiographic severity in subjects with advanced OA. Muscle activation, as quantified by EMG recordings, revealed neuromuscular compensations due to delayed quadriceps engagement and lessened co-contraction efficiency, usually not easily diagnosed during routine clinical assessments. All these objective parameters helped identify subtle functional decline even when patients themselves did not report any noticeable pain progression, thereby emphasizing the importance of continuous monitoring.

Home-based monitoring provided much richer insights into the participants' activities compared to short, discrete assessments that could have been obtained in-clinic. Over the course of the day, most subjects exhibited gait variability: morning stiffness was manifest as a lower cadence and reduced range of motion, whereas evening recordings often demonstrated asymmetry due to fatigue. In contrast, episodic hospital visits could not detect such temporal trends, which were demonstrated using this wearable system. These longitudinal data demonstrated insight into what to expect for the time course of symptoms to flare-up, such as increased step-to-step variability and irregular load distribution patterns preceding reported pain escalation. Such patterns can be considered early digital biomarkers, offering the potential for timely intervention, through exercises or modification of therapy. Another advantage of remote monitoring was improved patient compliance; users found the system to be comfortable and unobtrusive, making it easy to integrate smoothly into daily routines.

The correlation between sensor-derived metrics and clinical ground-truth assessments illustrated great agreement, confirming the accuracy and reliability of the proposed framework. The joint angle measurements showed high agreement with motion capture reference values, while pressure metrics strongly correlated with force plate data. EMG signal quality was consistently high due to the efficient preprocessing and noise-reducing techniques applied, which allowed the extraction of muscle activation events with high resolution of both timing and amplitude. It was observed that features from

combined modalities were able to discriminate among different OA severity levels significantly better than any single modality alone. Feature fusion in particular helped to improve the accuracy in the classification of monitoring outcomes, especially by fusing gait kinematics with loading distribution and EMG activation signatures. This demonstrates the efficiency of the system for long-term disease tracking, thus supplementing the use of traditional imaging-based assessments.

Model training yielded promising predictive performance regarding OA severity classification and symptom state estimation. Specifically, machine learning models showed high accuracy in differentiating between inter-day fluctuations, early worsening trends, and improved mobility after physiotherapy sessions. Feature importance analysis identified dynamic stiffness estimation, gait variability indices, and quadriceps activation delay as the top three most influential markers. These findings agree with established clinical knowledge but with improved sensitivity afforded by high-resolution, continuous measurements. The system's predictions of pain or stiffness episodes have reached high reliability, thus giving actionable insight to clinicians when optimizing treatment plans. Furthermore, patient-specific modeling demonstrated better predictive results compared to generalized models, which underlines the need for individualized monitoring when considering chronic musculoskeletal conditions. In summary, the findings presented confirm that wearable multisensor monitoring systems can offer a comprehensive, real-time evaluation of knee osteoarthritis progression beyond traditional diagnostic limitations. Integration of biomechanical and neuromuscular data offers a holistic view of joint function, allowing for earlier detection of deterioration and more responsive therapeutic management. This analysis demonstrates how sensor-based continuous assessment can significantly enhance clinical decision-making, support remote healthcare delivery, and empower patients with improved awareness of functional status. Its scalability, accuracy, and user acceptability make this system very promising for transforming knee OA diagnosis, monitoring, and rehabilitation both in the clinic and at home.

## CONCLUSION

Overall, the results of the present study show that wearable medical sensors for knee osteoarthritis assessment and monitoring are a transformational tool that surmounts many of the limitations of traditional diagnostic approaches. The continuous, high-resolution biomechanical and neuromuscular data obtained during natural daily movement provide in-depth insight into the

dynamism of joint function that is not achievable with the use of static imaging tools. Longitudinal measurements from IMUs, pressure sensors, EMG electrodes, and gyroscopes uncovered early deviations in gait mechanics, muscle activation, and the distribution of load, well before the development of clinically appreciable symptoms, indicative of the potential of this system for proactive disease management. Continuous monitoring thus allows not only very early detection of functional decline but also very personalized interventions, as sensor-derived metrics provide clinicians with the potential to individualize physiotherapy exercises, adjust orthotic devices, or modify treatment plans in line with real-world patient performance. Further, the demonstrated correlation between sensor-generated parameters and clinical ground truths confirms the reliability of the sensing system and its appropriateness for integration into routine healthcare workflows. Multimodal data fusion was of particular benefit, increasing overall diagnostic accuracy and allowing more accurate delineation between different OA severity levels compared to single-modality assessments. These results again stress the importance of including patient-specific sensor analytics, which demonstrated superior predictive capabilities and gave insights into each individual's unique disease course.

Besides diagnostic and prognostic gains, the study underlines the feasibility and practicality of remote OA monitoring through wearable technologies. The unobtrusive design and ease of use facilitated continuous adherence, hence allowing patients to participate in long-term monitoring without interference with daily activities. As further support for extended telemedicine efforts, this improves access for individuals who may face obstacles due to limited mobility or clinical resources. In addition, the detection of daily symptoms fluctuations allows clinicians to have a better view of real-life functional capacity, thus assisting in making more informed and evidence-based clinical decisions. This represents potential for scalability to broader deployment, benefiting diverse populations and healthcare settings. Advanced analytics, cloud-based processing, and personalized digital therapeutics, enabling automated alerts and adaptive rehabilitation programs, including predictive modeling of long-term disease course, can be integrated in the future. Conclusively, the study results spell out that wearable sensor-based monitoring is likely to be one of the important steps forward in knee osteoarthritis management, representing a powerful, innovative, patient-centered approach to early diagnosis, real-time assessment, and improvement in therapeutic outcomes.

## REFERENCES

1. R. Kaewsonthi, T. Sukpramote, N. Anantakool and W. Senavongse, "Gait Analysis and Fall Risk Assessment in Patients with Knee Osteoarthritis Using Different Measurement Tools," *2024 16th Biomedical Engineering International Conference (BMEiCON)*, Chon Buri, Thailand, 2024, pp. 1-5, doi: 10.1109/BMEiCON64021.2024.10896341.
2. C. Nichols, H. T. Crane, D. Ewart and O. T. Inan, "Combining Knee Acoustic Emissions, Patient-Reported Measures, and Machine Learning to Assess Osteoarthritis Severity," *2023 IEEE 19th International Conference on Body Sensor Networks (BSN)*, Boston, MA, USA, 2023, pp. 1-4, doi: 10.1109/BSN58485.2023.10331381.
3. K. Gangwar, R. S. C. Winter, F. M. Sabzevari, G. C.-Y. Chen, K. K.-M. Chan and K. Rambabu, "Monitoring Knee Health: Ultra-Wideband Radar Imaging for Early Detection of Osteoarthritis," *IEEE Transactions on Medical Imaging*, vol. 44, no. 8, pp. 3464-3475, Aug. 2025, doi: 10.1109/TMI.2025.3564521.
4. T. Guhan, R. Abiraj, M. Muneshwaran, V. S. Lakshmi and M. Suriyanarayanan, "Classification of Knee Osteoarthritis Using R-CNN Algorithm," *2025 International Conference on Multi-Agent Systems for Collaborative Intelligence (ICMS-Cl)*, Erode, India, 2025, pp. 1676-1680, doi: 10.1109/ICMS-Cl62561.2025.10894085.
5. S. D. Chandu, P. Revathi and N. A. S. Vinoth, "Discovering Knee Osteoarthritis Using CNN Enhanced with AlexNet," *2024 5th International Conference on Intelligent Communication Technologies and Virtual Mobile Networks (ICICV)*, Tirunelveli, India, 2024, pp. 134-141, doi: 10.1109/ICICV62344.2024.00028.
6. J. Al-Nabulsi, M. A.-S. Ahmad, B. Hasaneiah and F. AlZoubi, "Diagnosis of Knee Osteoarthritis Using Bioimpedance & Deep Learning," *2024 Second Jordanian International Biomedical Engineering Conference (JIBEC)*, Amman, Jordan, 2024, pp. 1-5, doi: 10.1109/JIBEC63210.2024.10931925.
7. G. Kitukale, N. A. Shelke, R. Agrawal, N. P. Singh and S. Quamara, "Predicting Knee Osteoarthritis using Deep Neural Network," *2024 IEEE 9th International Conference for Convergence in Technology (I2CT)*, Pune, India, 2024, pp. 1-6, doi: 10.1109/I2CT61223.2024.10543978.
8. C. J. Nichols, G. C. Özmen, K. Richardson, O. T. Inan and D. Ewart, "Classifying Pre-Radiographic Osteoarthritis of the Knee Using Wearable Acoustics Sensing at the Point of Care," *IEEE Sensors Journal*, vol. 23, no. 23, pp. 29619-29629, Dec. 2023, doi: 10.1109/JSEN.2023.3325153.
9. M. Shidore and R. Jalnekar, "Unveiling Risk Factors of Knee Osteoarthritis: A Novel Approach of Integrating Sample Survey Insights using Statistical and Machine Learning Model," *2023 3rd Asian Conference on Innovation in Technology (ASIANCON)*, Ravet, India, 2023, pp. 1-5, doi: 10.1109/ASIANCON58793.2023.10270497.
10. E. Panfilov, S. Saarakkala, M. T. Nieminen and A. Tiulpin, "End-to-End Prediction of Knee Osteoarthritis Progression With Multimodal Transformers," *IEEE Journal of Biomedical*

and *Health Informatics*, vol. 29, no. 9, pp. 6276-6286, Sept. 2025, doi: 10.1109/JBHI.2025.3536170.

- 11. M. Janotheepan, S. D. H. S. Wickramarathna, S. Bavathaaranie, M. Yanusha, K. Subahari and K. Thanusan, "Detection and Classification of Knee Osteoarthritis Using Convolutional Neural Network," *2024 4th International Conference on Advanced Research in Computing (ICARC)*, Belihuloya, Sri Lanka, 2024, pp. 132-136, doi: 10.1109/ICARC61713.2024.10499701.
- 12. L. Zhang et al., "Mechanism Design of Cable-driven Multi-functional Knee Osteoarthritis Rehabilitation Robot," *2024 30th International Conference on Mechatronics and Machine Vision in Practice (M2VIP)*, Leeds, U.K., 2024, pp. 1-6, doi: 10.1109/M2VIP62491.2024.10746128.
- 13. Y. Han, D. Shi, Y. Shao, Y. Feng, X. Ding and W. Zhang, "Design and Evaluation of a Soft Knee Exosuit for Reducing Knee Medial Compartment Load During Walking," *IEEE Transactions on Biomedical Engineering*, vol. 72, no. 5, pp. 1740-1749, May 2025, doi: 10.1109/TBME.2024.3520213.
- 14. Y. Nasser, M. El Hassouni, A. Chetouani and R. Jennane, "SWE-Net: Sliced-Wasserstein EfficientNet for Learning Discriminative Texture Features in Early Diagnosis of Knee Osteoarthritis," *2024 IEEE 13th International Conference on Image Processing Theory, Tools and Applications (IPTA)*, Rabat, Morocco, 2024, pp. 1-6, doi: 10.1109/IPTA62886.2024.10755821.
- 15. Y. Cai et al., "A method for gait phase detection in patients with knee osteoarthritis," *2025 International Conference on Mechatronics, Robotics, and Artificial Intelligence (MRAI)*, Jinan, China, 2025, pp. 474-479, doi: 10.1109/MRAI65197.2025.11135762.
- 16. A. Pandey and V. Kumar, "Enhancing Knee Osteoarthritis Severity Classification using Improved EfficientNet," *2023 10th IEEE Uttar Pradesh Section International Conference on Electrical, Electronics and Computer Engineering (UPCON)*, Gautam Buddha Nagar, India, 2023, pp. 1351-1356, doi: 10.1109/UPCON59197.2023.10434740.
- 17. R. Singh, N. Sharma, D. Upadhyay, S. Dangi and A. Singh, "Proposed CNN Model for the Classification of Knee Osteoarthritis," *2024 IEEE 3rd World Conference on Applied Intelligence and Computing (AIC)*, Gwalior, India, 2024, pp. 1052-1057, doi: 10.1109/AIC61668.2024.10730849.
- 18. N. Rameez, G. G. D. Dissanayake, G. A. C. Senindu, P. M. H. Liyanage and V. Embuldeniya, "Knee Osteoarthritis Detection Using Machine Learning on Pre-Processed X-ray Images and Treatment Recommendation," *2024 8th SLAAI International Conference on Artificial Intelligence (SLAAI-ICAI)*, Ratmalana, Sri Lanka, 2024, pp. 1-6, doi: 10.1109/SLAAI-ICAI63667.2024.10844957.
- 19. M. Arun, H. S. Krishnan, T. S. Kumar, N. Annamalai and C. Ravi Teja, "A Deep Learning Framework for Diagnosing and Predicting Knee Osteoarthritis Utilizing Convolutional Neural Networks," *2024 First International Conference on Software, Systems and Information Technology (SSITCON)*, Tumkur, India, 2024, pp. 1-6, doi: 10.1109/SSITCON62437.2024.10796117.
- 20. S. S. D. P and S. A. B, "Wearable Knee Cap Microwave Sensor for Osteoarthritis Detection," *2023 IEEE XXVIII International Seminar/Workshop on Direct and Inverse Problems of Electromagnetic and Acoustic Wave Theory (DIPED)*, Tbilisi, Georgia, 2023, pp. 171-175, doi: 10.1109/DIPED59408.2023.10269497.